



北京市津威康达医疗器械有限公司
Beijing Jinwei Kangda Medical Instrument Ltd.

Certificate of Analysis

Name	Cervical Collar (Soft)		Specification	M,L
The quantity of sample	4 Pieces		Lot No.	24101001
Model	KDZJ-JZ-016		Manufacturing Date	20241010
Inspection Reference	JINGFENGXIEBEI20140023		Test item	Product registration standard
Test results	Item	Standard	Actual results	Judgment for each
	Appearance	Should be no dirty ,no thread spliting, and no damaged , the both size of support bar and enhance belt are protected well.	Meet the standard	Qualified
	Support bar elasticity	Rigid support no deformation, elastic support flexible enough.		Qualified
	Velcro test	Velcro strong and firm, no disengagement or displacement		Qualified
	Lateral elasticity	1.2-1.7 times		Qualified
	Metal residue detection	Should pass the testing of table needle detection machine		Qualified
	Test condition	In the condition of 20-30 ℃ room temperature, use tape, needle detector and tensile testing machine test all the basic parameters.		
Inspection Conclusion	According to the test above all the test items are qualified.			



Checked by: Donglin Ge

Checked date: 2024/11/05

Company: Beijing Jinwei Kangda Medical Instrument Ltd.