

FOSHAN FEIYANG MEDICAL EQUIPMENT CO., LTD.
CERTIFICATE OF ANALYSIS

NO: FSYF-COA20250119001A

Product Name		STICK / BASTON		Lot.No.		FY20250119		QTY		1100PCS		Test QTY		1100PCS		Standard		GB/T 19545.1-2012 (ISO13485: 2016)	
Model No.		FY928L		Inspector		PENG HONGLIN		MFG Date		JAN.19,2025		checker		PENG LI		DATE		JAN.19,2025	
NO.		Name of item		Technology specification												Result		Single conclusion	
1		Appearance structure requirement		Appearance should be smooth, without irregularity , scars, distortion or deflection .												Compliant		Pass	
2		Dimension		Base size of stick should match the requirement of Form 1, limit deflection is no more than ±5cm.												Compliant		Pass	
3		Stability		The leg should be stable on the floor without swing.												Compliant		Pass	
4		Colour transfer		Plastic parts should have no colour transfer.												Compliant		Pass	
Test result		According to the test, all qualification reach the manufacture’s standard request GB/T 19545.1-2012 (ISO13485: 2016) Quality managing department: PAN QIU WEI Date:JAN.19,2025																	

