

Finished Product Inspection Report

NO: QA-202403053

Order No.		SM24010044		Lot No.		AS0224	
				Date of Production		2024.02	
Product Name		Latex-free Nonwoven Cohesive Bandage		Specification (cm*m)		2.5/5/7.5/10*4.5	
Color		Beige/White/Light Blue/Green//Pink/Orange		Quantity (cartons)		97	
Quantity (rolls)		22080		Tested (rolls)		20	
Items of inspection		Acceptable		Instrument		Test result /Nonconformity description	
Products	Basic specs	Appearance		clean, no clearly eyewinker	Eye survey	clean	conforming
		Full diameter(mm)		54±3	Vernier caliper	53.29	conforming
		Stretch Length(m)		≥4.5	Tape measure	4.92	conforming
		Width(mm)		25/50/75/100±1	Ruler	24.9/49.9/74.8/99.9	conforming
	Performance specs	Adhesion to itself(N/inch)		≥1.25	Tensile Strength Machine	2.81	conforming
		Elasticity		1: (2.1±0.3)	Ruler	1:2.18	conforming
Package	Carton	Appearance		clean, no clearly eyewinker	Eye survey	clean	conforming
		Words/pattern		conforming	Eye survey	correct	conforming
	Core	Appearance		clean, no clearly eyewinker	Eye survey	clean	conforming
		Words/pattern		conforming	Eye survey	correct	conforming
Result: conforming							

Inspector

chen

2024.3.15

Quality Manager

X. J. Li

2024.3.15

JSCP-QA-R09A

质量专用章

Jiangsu Yizheng Carepro Medical Instruments Co., Ltd.