

FOSHAN FEIYANG MEDICAL EQUIPMENT CO., LTD.
CERTIFICATE OF ANALYSIS

NO: FSFY-COA20230602001A

Product Name		CRUTCH / MULETA	Lot.No.	FY20230525	QTY	2200PCS	Test QTY	2200PCS	Standard	GB/T 19545.2-2009 (ISO13485: 2016)
Model No.		FY925L(L)	Inspector	ZOU MENG XIA	MFG Date	June.2,2023	checker	PENG LI	DATE	June.2,2023
NO.	Name of item	Technology specification							Result	Single conclusion
1	Appearance structure requirement	Appearance should be smooth, without irregularity , scars, distortion or deflection,without color fade,screw without loosen							Compliant	Pass
2	Dimension	size of crutch should match the requirement of Form 1, limit deflection is no more than ±5cm.							Compliant	Pass
3	Stability	leg should be stable on the floor without swing.							Compliant	Pass
4	Colour transfer	Plastic parts should have no colour transfer.							Compliant	Pass
Test result		According to the test, all qualification reach the manufacture’s standard request GB/T 19545.2-2009 (ISO13485: 2016). <div>Quality managing department: PAN QIU WEI Date: June.2,2023</div>								

