

FOSHAN FEIYANG MEDICAL EQUIPMENT CO., LTD.

CERTIFICATE OF ANALYSIS

NO: FSYF-COA20220310001A

Product Name		WALKER/ ANDADOR	Lot.No.	FY20211125	QTY	800PCS	Test QTY	800PCS	Standard	GB/T 14728.1-2006 (ISO13485: 2016)
Model No.		FY961L	Inspector	PENG LI	MFG Date	March.10,2022	checker	PENG LI	DATE	March.10,2022
NO.	Name of item	Technology specification							Result	Single conclusion
1	Appearance structure requirement	Appearance should be smooth, without irregularity , scars, distortion or deflection .							Compliant	Pass
2	Dimension	Base size of walker should match the requirement of Form 1, limit deflection is no more than ±5cm.							Compliant	Pass
3	Stability	Four legs should be stable on the floor without swing.							Compliant	Pass
4	Colour transfer	Plastic parts should have no colour transfer.							Compliant	Pass
Test result		According to the test, all qualification reach the manufacture’s standard request GB/T 14728.1-2006 (ISO13485: 2016) . Quality managing department: PAN QIU WEI Date:March.10,2022								

