

# FOSHAN FEIYANG MEDICAL EQUIPMENT CO., LTD.

## CERTIFICATE OF ANALYSIS

NO: FSYF-COA20211118002A

Product Name		STICK / BASTON		Lot.No.		FY20211115		QTY		360PCS		Test QTY		360PCS		Standard		GB/T 19545.1-2012 (ISO13485: 2016)	
Model No.		FY936L		Inspector		ZOU MENG XIA		MFG Date		Nov.18,2021		checker		PENG LI		DATE		Nov.18,2021	
NO.		Name of item		Technology specification												Result		Single conclusion	
1		Appearance structure requirement		Appearance should be smooth, without irregularity , scars, distortion or deflection .												Compliant		Pass	
2		Dimension		Base size of commode should match the requirement of Form 1, limit deflection is no more than ±5cm.												Compliant		Pass	
3		Stability		Four legs should be stable on the floor without swing.												Compliant		Pass	
4		Colour transfer		Plastic parts should have no colour transfer.												Compliant		Pass	
Test result		According to the test, all qualification reach the manufacture’s standard request GB/T 19545.1-2012 (ISO13485: 2016) <div>Quality managing department: PAN QIU WEI    Date:Nov.18,2021</div>																	

