

电子血压计成品检验报告
FQC Report of Digital blood pressure minitor

Manufacturing Date: 2025/4/28
Expiration date: N/A

检验依据 Inspection basis: 电子血压计成品检验工艺 NTLD-XYJ-SIP101 Final inspection process of BPM								
产品型号 Product Model : 2522			报检数量 Quantity declared : 1000			产品批号 Lot No.: 25YF1B0271		
温度 Temperature: 25°C			湿度 Humidity: 60%			检验人 Inspected by: 李平		
产品编号 Product No.: 25220001-252201000			检验数量 Quantity inspected: 13/50			检验日期 Inspection date: 2025-6-30		
订单号 NO: 25YF1B027R1			版本 Version: 26			机型 Model: B3		
序号 No.	检验项目 Inspection items	检测设备 / 设备编号 Inspection equipment/ Equipment serial number	检验水平 Testing level	AQL 值 AQL value	抽样数 Sample	不良数 RC	判定 Judge	备注 Note
1	压力示值误差 ≤ ±2 mmHg Static Pressure Deviation ≤ ± 2 mmHg	精密压力表 PP01-11 Precise Manometer	S-3	C=20	13	0	ACC	
2	放气速度检测 (针对降压机型) Leakage Rate (for lower Pressure type)	气流测试仪 Air Flow Meter	/	/	/	/	/	
3	气密测试 (针对升压机型) ≤ 3 air tightness test (for upper pressure type) ≤ 3	气流测试仪 AF02-02 Air Flow Meter	S-3	C=20	13	0	ACC	
4	低电压提示 Low Voltage Indication	万用表 FM02-23 Multimeter	S-3	C=20	13	0	ACC	
5	压力上限提示, 可泄放 after the upper limit of pressure, it can be released	目测 Visual inspection	/	/	/	/	/	
	快速放气压力从 260 到 15mmHg ≤ 10s The time is within 10S from 260mmHg to 15mmHg	血压模拟器 / 秒表 Cufflink/Stopwatch	/	/	/	/	/	
6	3min 无操作自动关机 Auto switched off in 3 min when no operate	秒表 H101-14 Stopwatch	S-1	C=20	5	0	ACC	
7	脉搏测量 ± 5% Pulse Measurement ± 5%	血压计模拟器 Cufflink FB5-08	S-3	C=20	13	0	ACC	
8	一般功能检查 General Function inspection	目测 Visual inspection	S-3	C=20	13	0	ACC	
9	配件检测 Accessories test	目测/精密压力表 PP01-11 Visual inspection	S-1	C=20	5	0	ACC	
10	外观检查 Appearance	目测 Visual inspection	I	0.4/1.0	50	0	ACC	
11	包装检查 Packing inspection	目测 Visual inspection	I	0.4/1.0	50	0	ACC	
综合判定 Conclusion: 合格								
核准 Approved: 周美玲			日期 Date: 2025-6-30					

随附检测记录 (一)(二) Inspection record is attached.

文件编号 Document No.: NTLD-QS-209-F08

版本 d

实施日期: 2022.11.08

上海强生国际贸易有限公司
SHANGHAI HONSUN INTERNATIONAL TRADE CO., LTD

朱文军

检测记录(二)
Inspection Record(2)

检测项目 产品编号 Item Serial number	全显检测 All display test	机型检测 Model test	IHB检测 IHB Test	kPa与mmHg切换功能 Switch function between kPa and mmHg	测量结果显示功能 Measurement result display function	摄氏度与华氏切换 Celsius and Fahrenheit switch	时间功能调整 Time function adjustment	语言选择及音量调节功能 Language selection and volume adjustment function	记忆调出功能 Memory callout function	记忆存入功能 Memory storage function	闹钟设置 Alarm clock settings	三色背光检测 Three-color or backlight test	条码核对 Barcode check	蓝牙开关调整 Bluetooth switch adjustment	蓝牙连接 Bluetooth connection	蓝牙同步测试 Bluetooth synchronization test	判定 Judge
2552200017	✓	✓	✓	/	✓	/	/	/	✓	✓	/	/	/	/	/	/	✓
00014	✓	✓	✓	/	✓	/	/	/	✓	✓	/	/	/	/	/	/	✓
00032	✓	✓	✓	/	✓	/	/	/	✓	✓	/	/	/	/	/	/	✓
00033	✓	✓	✓	/	✓	/	/	/	✓	✓	/	/	/	/	/	/	✓
00206	✓	✓	✓	/	✓	/	/	/	✓	✓	/	/	/	/	/	/	✓
00211	✓	✓	✓	/	✓	/	/	/	✓	✓	/	/	/	/	/	/	✓
00437	✓	✓	✓	/	✓	/	/	/	✓	✓	/	/	/	/	/	/	✓
00436	✓	✓	✓	/	✓	/	/	/	✓	✓	/	/	/	/	/	/	✓
00611	✓	✓	✓	/	✓	/	/	/	✓	✓	/	/	/	/	/	/	✓
00614	✓	✓	✓	/	✓	/	/	/	✓	✓	/	/	/	/	/	/	✓
00622	✓	✓	✓	/	✓	/	/	/	✓	✓	/	/	/	/	/	/	✓
00831	✓	✓	✓	/	✓	/	/	/	✓	✓	/	/	/	/	/	/	✓
00834	✓	✓	✓	/	✓	/	/	/	✓	✓	/	/	/	/	/	/	✓

检测项目检测合格标识: ✓, 不合格注明不合格问题