

LATEX SURGICAL GLOVES

Quality Standards

- ◆ Meets or exceeds the following standards: ASTM D3577-91 (USA), EN 455 (EEC), AS 40 (Australia), FDA, CE
- ◆ Quality sample in accordance with MIL STD 105D
- ◆ ISO 9002 Certified Manufacturing
- ◆ Meets the Viral Penetration Test (ASTM F1671-97)



Sizes	Item#	Glove Length	Palm width +/-10mm	Thickness		
				Cuff	Palm	Finger
6	030-6.0	300	76	0.14	0.16	0.18
6 1/2	030-6.5	300	83	0.14	0.16	0.18
7	030-7.0	300	89	0.14	0.16	0.18
7 1/2	030-7.5	300	95	0.14	0.16	0.18
8	030-8.0	300	102	0.14	0.16	0.18
8 1/2	030-8.5	300	108	0.14	0.16	0.18

Physical Properties

Before Aging: Accelerated Aging:
7 days @ 70°C



■ Ideal Medical's Tensile Strength (Mpa)
■ ASTM Minimum Requirement

Physical Properties

Before Aging: Accelerated Aging:
7 days @ 70°C



■ Ideal Medical's Ultimate Elongation
■ ASTM Minimum Requirement