

CERTIFICATE OF ANALYSIS

NO: FSFY-COA20211019001A

Product Name		CRUTCH / MULETA		Lot.No.	FY20211025	QTY	4000PCS	Test QTY	4000PCS	Standard	GB/T 19545.2-2009 (ISO13485: 2016)
Model No.		FY925L(M)		Inspector	ZOU MENG XIA	MFG Date	Oct.18-19,2021	checker	PENG LI	DATE	Oct.19,2021
NO.	Name of item	Technology specification								Result	Single conclusion
1	Appearance structure requirement	Appearance should be smooth, without irregularity , scars, distortion or deflection,without color fade,screw without loosen								Compliant	Pass
2	Dimension	size of crutch should match the requirement of Form 1, limit deflection is no more than ± 5 cm.								Compliant	Pass
3	Stability	legs should be stable on the floor without swing.								Compliant	Pass
4	Colour transfer	Plastic parts should have no colour transfer.								Compliant	Pass
Test result		According to the test, all qualification reach the manufacture's standard request GB/T 19545.2-2009 (ISO13485: 2016). Quality managing department: PAN QIU WEI Date: Oct.19,2021									