

## Certificate of Analysis

|              |                                                                                                                                                                                                                                               |            |                |
|--------------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|------------|----------------|
| Product Name | Master Cardiology Stethoscope                                                                                                                                                                                                                 | Model      | Deluxe         |
| QTY          | 500 pcs                                                                                                                                                                                                                                       | Lot Number | 20240715       |
| MFG.Date     | 20240715                                                                                                                                                                                                                                      | Test Date  | 2024.09.15     |
| No.          | Test Item & Requirement                                                                                                                                                                                                                       | Result     | Single Verdict |
| 1            | The sound sending of stethoscope shall clear.                                                                                                                                                                                                 | Conformity | Qualified      |
| 2            | The earring's elasticity shall feasible, when 2 earplugs have been undrawn to 300mm distance, after comeback, the distortion distance shall less than 10mm .                                                                                  | Conformity | Qualified      |
| 3            | The earring bounce shall moderate.                                                                                                                                                                                                            | Conformity | Qualified      |
| 4            | Three-way catheter's surface shall slick and clean, without air bubble and distortion disfigurements etc.                                                                                                                                     | Conformity | Qualified      |
| 5            | Diaphragm shall clean, without nick, pock and besmirch disfigurements etc. Diaphragm shall not become flexible.                                                                                                                               | Conformity | Conformity     |
| 6            | All parts of stethoscope shall fixed, not become flexible.                                                                                                                                                                                    | Conformity | Conformity     |
| 7            | Each parts' appearance shall symmetry, without crack, sunken and plating break off etc. Plating parts' appearance shall brightness, surface roughness degree Ra0.8, without yellow uncovered, pocked, break off and nicks disfigurements etc. | Conformity | Conformity     |

Inspected by / Date: Li Kejie / 2024.09.15

Approved by / Date: He Ming / 2024.09.15

