

HUAIAN WANJIA MEDICAL DEVICE CO.,LTD.

TEL:0086-517-84519182

FAX:0086-517-83765785

ANALYSIS REPORTS OF STERILE CARBON SURGICAL BLADE

Merchandise name	Sterile surgical carbon steel blade 12#	Standard
Q NUMBER:E-19225B	LOT:20190816	YY0293,BS2982;Sampling method:GB/T2828
Item of specification	Standard require	Conclusion of inspection
Hardness	$\geq 800\text{HV}$	825 830 847 832 824 822 848 810 806 819 805 815 826
Sharpness	$\leq 0.80\text{N}$	0.37 0.59 0.53 0.54 0.53 0.55 0.58 0.61 0.59 0.51 0.45 0.59 0.49
Elasticity	No clear distortion	Passed
Surface	The luster should be level ,smooth surface, even blade body ,no contamination,	Passed
Packing mark	All labels have blade size ,Lot No., sterilized method ,expiry date and manufacturer	Passed
Sterilized test	Sterile	Sterile

Conclusion: This Lot of Products is qualified

QUALITY DIRECTOR: Yongchao Cheng

DATE:NOV 20,2019

