

FOSHAN FEIYANG MEDICAL EQUIPMENT CO., LTD.

CERTIFICATE OF ANALYSIS

NO: FSFY-COA20230607001A

Product Name		WALKER/ ANDADOR		Lot.No.	FY20230525	QTY	600PCS	Test QTY	600PCS	Standard	GB/T 14728.1-2006 (ISO13485: 2016)
Model No.		FY913L		Inspector	ZOU MENG XIA	MFG Date	June.7,2023	checker	PENG LI	DATE	June.7,2023
NO.	Name of item		Technology specification							Result	Single conclusion
1	Appearance structure requirement		Appearance should be smooth, without irregularity , scars, distortion or deflection .							Compliant	Pass
2	Dimension		Base size of commode should match the requirement of Form 1, limit deflection is no more than ±5cm.							Compliant	Pass
3	Stability		Four legs should be stable on the floor without swing.							Compliant	Pass
4	Weight capacity		Walker should have enough strength. It should be bear 100kg weight capacity.							Compliant	Pass
5	Colour transfer		Plastic parts should have no colour transfer.							Compliant	Pass
Test result		According to the test, all qualification reach the manufacture’s standard request GB/T 14728.1-2006 (ISO13485: 2016). Quality managing department: PAN QIU WEI Date: June.7,2023									

