FOSHAN FEIYANG MEDICAL EQUIPMENT CO., LTD. <u>CERTIFICATE OF ANALYSIS</u>

NO: FSFY-COA2025032002A

Produ Nam	roduct WALk Name ANDA			Lot.No.	FY20250329	QTY	280PCS	Test QTY	280PCS	Standard	GB/T 14728.1-2006 (ISO13485: 2016)
Model No.		FY913L		Inspector	Peng honglin	MFG Date	Mar.29,2025	checker	PENG LI	DATE	Mar.29,2025
NO.	Name of item		Technology specification							Result	Single conclusion
1	Appearance structure requirement		Appearance should be smooth, without irregularity, scars, distortion or deflection.							Compliant	Pass
2	Dimension		Base size of walker should match the requirement of Form 1, limit deflection is no more than ± 5 cm.							Compliant	Pass
3	Stability		Four legs should be stable on the floor without swing.							Compliant	Pass
4	Weight capacity		Walker should have enough strength. It should be bear 100kg weight capacity.							Compliant	Pass
5	Colour transfer		Plastic parts should have no colour transfer.							Compliant	Pass
Test result According to the test, all qualification reach the manufacture's standard request GB/T 14728.1-2006 (ISO13485: 2016). Quality managing department: PAN QIU WEI Date:										EI Date: Mar.29,2025	