

FOSHAN FEIYANG MEDICAL EQUIPMENT CO., LTD.

CERTIFICATE OF ANALYSIS

NO: FY-COA20220618001A

Product Name		WALKER/ ANDADOR	Lot.No.	FY20220610	QTY	200PCS	Test QTY	200PCS	Standard	GB/T 14728.1-2006 (ISO13485: 2016)
Model No.		FY913L(S)	Inspector	ZOU MENG XIA	MFG Date	June.18,2022	checker	PENG LI	DATE	June.18,2022
NO.	Name of item	Technology specification							Result	Single conclusion
1	Appearance structure requirement	Appearance should be smooth, without irregularity , scars, distortion or deflection .							Compliant	Pass
2	Dimension	Base size of commode should match the requirement of Form 1, limit deflection is no more than ± 5cm.							Compliant	Pass
3	Stability	Four legs should be stable on the floor without swing.							Compliant	Pass
4	Colour transfer	Plastic parts should have no colour transfer.							Compliant	Pass
Test result		According to the test, all qualification reach the manufacture’s standard request GB/T 14728.1-2006 (ISO13485: 2016) . Quality managing department: PAN QIU WEI Date: June.18,2022								

