

Certificate of Analysis

Product Name		Microscope Slides	Model	7105
Total Quantity		15000Boxes	LOT	20240715
Mfg. Date		20240715	Test Date	2024.09.25
No	Inspection Item	Inspecting Standard	Inspecting Tool	Result
1	Material	Made from soda glass	Visual	Conformity
2	Surface	colorless and transparent,good light transmittance, smooth,no adhesion phenomenon.Clean surface,should not have fog, water traces and fingerprints and other defects. Should not be visible visual windbreak pits, granules, stones, scratches, fractures, and other defects. There are no bubbles, stripes, inclusions, spots, scratches and other defects in the central area.	Visual	Conformity
3	Edge	Light green, polisged edges should be neat, no defect.	Visual	Conformity
4	Corner	45 degree	Visual	Conformity
5	Size	25.4x76.2mm	Vernier Caliper	Conformity
6	Thickness	1.0-1.2mm	Vernier Caliper	Conformity
7	Packaging	72pcs/box,with tissue paper interleaved	Visual	Conformity
8	Inside Packing	Transparent PE bag vacuum-packed	Visual	Conformity
9	Outer Packing	The outer printing is complete and the logo is clear. The outer carton includes the product name, specification, box gauge, gross net weight	Visual	Conformity
Conclusion: The product meets the requirements and is judged to be qualified				

Tester: Lin yan

Checker: Sun Juping

