

FOSHAN FEIYANG MEDICAL EQUIPMENT CO., LTD.

CERTIFICATE OF ANALYSIS

NO: FY-COA20220614001A

| | | | | | | | | | | |
|--------------|----------------------------------|--|-----------|--------------|----------|--------------|----------|---------|-----------|------------------------------------|
| Product Name | | STICK / BASTON | Lot.No. | FY20220610 | QTY | 200PCS | Test QTY | 200PCS | Standard | GB/T 19545.1-2012 (ISO13485: 2016) |
| Model No. | | FY929L | Inspector | ZOU MENG XIA | MFG Date | June.14,2022 | checker | PENG LI | DATE | June.14,2022 |
| NO. | Name of item | Technology specification | | | | | | | Result | Single conclusion |
| 1 | Appearance structure requirement | Appearance should be smooth, without irregularity , scars, distortion or deflection . | | | | | | | Compliant | Pass |
| 2 | Dimension | Base size of commode should match the requirement of Form 1, limit deflection is no more than ± 5cm. | | | | | | | Compliant | Pass |
| 3 | Stability | Four legs should be stable on the floor without swing. | | | | | | | Compliant | Pass |
| 4 | Colour transfer | Plastic parts should have no colour transfer. | | | | | | | Compliant | Pass |
| Test result | | According to the test, all qualification reach the manufacture’s standard request GB/T 19545.1-2012 (ISO13485: 2016) | | | | | | | | |
| | | Quality managing department: PAN QIU WEI Date:June.14,2022 | | | | | | | | |

