FOSHAN FEIYANG MEDICAL EQUIPMENT CO., LTD. <u>CERTIFICATE OF ANALYSIS</u>

NO: FSFY-COA20211013001A

Product Name		STICK / BASTON		Lot.No.	FY20211025	QTY	1000PCS	Test QTY	1000PCS	Standard	GB/T 19545.1-2012 (ISO13485: 2016)
Model No.		. FY937L		Inspector	ZOU MENGXIA	MFG Date	Oct.13,2021	checker	PENG LI	DATE	Oct.13,2021
NO.	Name of item			Technology specification							Single conclusion
1	Appearance structure requirement		Appearance should be smooth, without irregularity, scars, distortion or deflection, without color fade, screw without loosen							Compliant	Pass
2	Dimension		size of crutch should match the requirement of Form 1, limit deflection is no more than ± 5 cm.							Compliant	Pass
3	Stability		legs should be stable on the floor without swing.							Compliant	Pass
4	Colour transfer		Plastic parts should have no colour transfer.							Compliant	Pass
Те	est	According to the test, all qualification reach the manufacture's standard request GB/T 19545.1-2012 (ISO13485: 2016)									
res	ult	Quality managing department: PAN QIU WEI Date: Oct.13,2021									