FOSHAN FEIYANG MEDICAL EQUIPMENT CO., LTD. <u>CERTIFICATE OF ANALYSIS</u>

NO: FSFY-COA20211019002A

Product Name		CRUTCH / MULETA		Lot.No.	FY20211025	QTY	400PCS	Test QTY	400PCS	Standard	GB/T 19545.2-2009 (ISO13485: 2016)
Model	No.	FY925L(S)		Inspector	ZOU MENG XIA	MFG Date	Oct.19,2021	checker	PENG LI	DATE	Oct.19,2021
NO.	Name of item		Technology specification							Result	Single conclusion
1	Appearance structure requirement		Appearance should be smooth, without irregularity, scars, distortion or deflection, without color fade, screw without loosen							Compliant	Pass
2	Dimension		size of crutch should match the requirement of Form 1, limit deflection is no more than ± 5 cm.							Compliant	Pass
3	Stability		legs should be stable on the floor without swing.							Compliant	Pass
4	Colour transfer		Plastic parts should have no colour transfer.							Compliant	Pass
Тє	est	According to the test, all qualification reach the manufacture's standard request GB/T 19545.2-2009 (ISO13485: 2016).									
res	ult	Quality managing department: PAN QIU WEI Date: Oct.19,2021									