## FOSHAN FEIYANG MEDICAL EQUIPMENT CO., LTD. CERTIFICATE OF ANALYSIS

NO: FSFY-COA20231203001A

Product Name		WALKER/ ANDADOR		Lot.No.	FY20231130	QTY	600PCS	Test QTY	600PCS	Standard	GB/T 14728.1-2006 (ISO13485: 2016)
Model No.		FY963L		Inspector	ZOU MENG XIA	MFG Date	NOV.30,2023	checker	PENG LI	DATE	Dec.3,2023
NO.	Name of item			Technology specification							Single conclusion
1	Appearance structure requirement		Appearance should be smooth, without irregularity, scars, distortion or deflection.							Compliant	Pass
2	Dimension		Base size of commode should match the requirement of Form 1, limit deflection is no more than $\pm 5 \text{cm}$ .							Compliant	Pass
3	Stability		Four legs should be stable on the floor without swing.							Compliant	Pass
4	Weight capacity		Walker should have enough strength. It should be bear 100kg weight capacity.							Compliant	Pass
5	Colour transfer		Plastic parts should have no colour transfer.							Compliant	Pass
Test result  According to the test, all qualification reach the manufacture's standard request GB/T 14728.1-2006 (ISO13485: 2016).  Quality managing department: PAN QIU W									WEI Date: Dec.3,2023		