

TOP GLOVE SDN BHD (MANUFACTURER)

Certificate of Analysis

Customer ALESSA S.A.
Brand TG [ALESSA SA, MEDO CW] DIM D
Sales Order 5400044359
Customer PO -
Type of Glove NITRILE POWDER FREE ONLINE SINGLE CHLORINATED GLOVE (Finger Textured, Black)
Glove Grade CW040
AQL Required 1.5
Date of release 30.03.2026
Reference Standard The above consignment of goods have been inspected against Top Glove Standard where samples selected at random using Single Sampling Plans for Normal Inspection of ISO 2859-1.
Declared Quantity, Size, Lot No., Manufacturing Date, Expiry Date

COA4025/5400044359/000

Size	Quantity (Carton)	Quantity (Pieces)	Lot No.	Manufacturing Date	Expiry Date
M	400	400,000	225044359NCZA	2026-02-21	2031-02-20
L	700	700,000	225044359NCZA	2026-02-21	2031-02-20
TOTAL	1,100	1,100,000			

Characteristics	Size	Unit	Inspection Level	Acceptance	Result
Freedom From Holes					
	M	PC	GI	AQL 1.5	PASS
	L	PC	GI	AQL 1.5	PASS

Characteristics	Size	Unit	Inspection Level	Acceptance	Result
Major Visual Defect					
	M	PC	GI	AQL 2.5	PASS
	L	PC	GI	AQL 2.5	PASS

Characteristics	Size	Unit	Inspection Level	Acceptance	Result
Minor Visual Defect					
	M	PC	GI	AQL 4.0	PASS
	L	PC	GI	AQL 4.0	PASS

Characteristics	Size	Unit	Sample Size	Non-Compliant	Acceptance	Min. Value	Max. Value	Result
Glove Length						S2_4.0MV		
	M	mm	13	0	1	240	243	PASS
	L	mm	13	0	1	243	245	PASS

Characteristics	Size	Unit	Sample Size	Non-Compliant	Acceptance	Min. Value	Max. Value	Result
Glove Palm Width						S2_4.0MV		
	M	mm	13	0	1	94	96	PASS
	L	mm	13	0	1	104	106	PASS

Characteristics	Size	Unit	Sample Size	Non-Compliant	Acceptance	Min. Value	Max. Value	Result
Fingertip Thickness					S2_4.0MV			
	M	mm	13	0	1	0.12	0.14	PASS
	L	mm	13	0	1	0.12	0.13	PASS

Characteristics	Size	Unit	Sample Size	Non-Compliant	Acceptance	Min. Value	Max. Value	Result
Palm Thickness					S2_4.0MV			
	M	mm	13	0	1	0.07	0.08	PASS
	L	mm	13	0	1	0.06	0.07	PASS

ASTM D6319-19 Requirement:

Length (mm) :	Width (mm) :	Thickness (Fingertip & Palm) :
XS, S : Min. 220	XS : 70 ± 10	Min. 0.05mm
M, L, XL, XXL : Min. 230	S : 80 ± 10	
	M : 95 ± 10	
	L : 110 ± 10	
	XL : 120 ± 10	
	XXL : 130 ± 10	

Characteristics	Size	Unit	Sample Size	Non-Compliant	Acceptance	Min. Value	Max. Value	Result
Tensile Strength Before Aging					S2_4.0MV			
	M	MPa	13	0	1	24.3	40.9	PASS
	L	MPa	13	0	1	17.1	31.5	PASS

Characteristics	Size	Unit	Sample Size	Non-Compliant	Acceptance	Min. Value	Max. Value	Result
Tensile Strength After Aging					S2_4.0MV			
	M	MPa	13	0	1	22.1	36.4	PASS
	L	MPa	13	0	1	18.1	41.3	PASS

Characteristics	Size	Unit	Sample Size	Non-Compliant	Acceptance	Min. Value	Max. Value	Result
Elongation Before Aging					S2_4.0MV			
	M	%	13	0	1	503.0	638.0	PASS
	L	%	13	0	1	515.0	610.0	PASS

Characteristics	Size	Unit	Sample Size	Non-Compliant	Acceptance	Min. Value	Max. Value	Result
Elongation After Aging					S2_4.0MV			
	M	%	13	0	1	467.0	580.0	PASS
	L	%	13	0	1	450.0	576.0	PASS

ASTM D6319-19 Requirement:

Before Aging :	After Aging :
Tensile Strength : Min. 14Mpa	Tensile Strength : Min. 14Mpa
Elongation at Break : Min. 500%	Elongation at Break: Min. 400%

Characteristics	Size	Unit	Sample Size	Mean Value	Result
Powder Free Residue (Powderfree Glove)			FIXTR_5		
	M	mg/glo	5	0.63	PASS
	L	mg/glo	5	0.21	PASS

Requirement : Max. 2mg/glove

Conclusion

We hereby certify that glove consignment of goods were determined to meet the acceptable limit of the specifications as referring to the above findings of randomly selected samples.

Prepared By
AISWARIA A/P. RAMALINGAM
EXECUTIVE, QA

Approved By
SIM PECK KUNG
SNR. MANAGER, QA

This is a computer generated test report, no signature is required.